

For Any Technical Questions, Please Contact at info@nodesus.com

Product Name	operation temperature	stabilization time(sec)	Measuring range
PH test pen (Accuracy 0.1)	0-60°C	25°C ≤ 25S	0-14PH at 0-100°C
PH test pen (Accuracy 0.01)	0-60°C	25°C ≤ 25S	0-14PH at 0-100°C
PH test pen (Accuracy 0.1 External electrode)	0-60°C	25°C ≤ 25S	0-14PH at 0-100°C
PH test pen (Accuracy 0.01 External electrode)	0-60°C	25°C ≤ 25S	0-14PH at 0-100°C